

# REVEAL® Q+ FOR FUMONISIN

DOWNLOAD AND READ KIT INSTRUCTIONS COMPLETELY BEFORE PERFORMING TEST.



## Materials Provided:

- 25 Reveal® Q+ for Fumonisin test strips
- 25 Red sample dilution cups
- 25 Clear sample cups
- 01 Bottle of sample diluent

**Product Number: 8885**

**Threshold: 0.3–6 ppm**


**Testing time: 6 minutes**

Call 800.234.5333 to order or visit [NEOGEN.com](http://NEOGEN.com)

## Procedure



1 Prepare by entering the QR code into the Raptor® Solo reader.



2 Obtain a representative sample. Grind and weigh out a 10 g sample.




3 Add 50 mL of 65% ethanol to the sample.



4 Shake vigorously for 3 minutes, or blend for 1 minute.



5 Allow to settle, then filter or centrifuge sample.



6 Add 200 µL sample extract to the red dilution cup.



7 Add 400 µL of sample diluent to the red dilution cup mix up and down 5 times.




8 Insert strip into the Raptor cartridge.



9 Insert cartridge into the Raptor Solo reader.



10 Add 400 µL sample from the red dilution cup to the back of the Raptor cartridge. The reader will start automatically,



11 Results shown on the Raptor Solo reader. Note: samples over 6 ppm must be diluted and re-tested.



# Reveal<sup>®</sup> Q+ for Fumonisin

Product Number: 8885

## Ordering Information

- 8885 Reveal Q+  
for Fumonisin
- 9696 Raptor Solo platform
- 9681 Raptor cartridges



## Materials Recommended, Not Provided

NEOGEN #	Item Description
9428, 9428B	Sample collection cups with lids, 125 mL
9330	Centrifuge, mini
9172	Microcentrifuge tubes
9448, 9447	Dispensing pump or graduated cylinder
9401, 9453	Agri-grind grinder or equivalent
9427	Scale capable of weighing 5–50 g ± 0.1 g
9426, 9452	Timer
8073, 8074	65% ethanol solution
9421, 9421B	Sample collection tubes with caps
9420, 9429, 9519	Filter syringes, Whatman #4 filter paper or equivalent
9475	Reveal sample cup rack
9463	Pipettor, 100–1000 µL
9693	Pipettor, 400 µL
9464, 9487	Pipette tips, 100–1000 µL
—	Distilled or deionized water

## Test with Confidence

NEOGEN<sup>®</sup>'s Reveal<sup>®</sup> Q+ for Fumonisin, an easy-to-use quantitative lateral flow test that provides unparalleled accuracy in only six minutes.

The Raptor<sup>®</sup> reader allows the user to simply add sample and walk away — the instrument controls the volume, temperature and timing of your assay. The system allows you to run three samples at one time.

- **Simple** procedure — just add sample
- **Fast** time to result
- **Eliminate** operator errors

